

# Asnis<sup>®</sup> & Fixos<sup>™</sup> Screw Systems

stryker<sup>®</sup>



Rearfoot



Midfoot



Forefoot

The Stryker screw systems offer multiple configurations to address a variety of Forefoot, Midfoot and Rearfoot bone repair procedures:

- Ø2.5, 3.5 & 4.0mm Headless Cannulated Compression Screws
- Ø2.0 & 2.7mm Twist-Off Screws
- Ø2.0, 3.0, 4.0, 5.0, 6.5 & 8.0mm Cannulated Screws



Foot & Ankle

## Related Products

### External Fixation Systems



Hoffmann 3

### Plating Systems



VariAx Foot & Fibula Anchorage

### IM Nail



T2 Ankle

### Biologics



Vitoss

# Asnis & Fixos

## Screw Systems

stryker®

### Asnis



#### Asnis Micro

- Low Profile Screw Head
- Titanium
- Shaft and Core Diameter Equal
- Reverse Cutting Flute To facilitate removal
- Self-Drilling/Tapping Design
- Torx Screwdriver T6, T10
- Guide Wire Diameter 0.8mm & 1.2mm



#### Asnis III

- Low Profile Screw Head
- Titanium / Stainless Steel
- Shaft and Core Diameter Equal
- Thread Choice Implants available with both partial and fully threaded options
- Reverse Cutting Flute To facilitate removal
- Self-Drilling/Tapping Design
- Hex Screwdriver, 2.5, 3.5, 5.0mm
- Large Diameter Guide Wires 1.4, 2.0, 3.2mm To allow for precise screw placement, the threaded tip engages bone for better bone purchase

4.0, 5.0, 6.5, 8.0mm

### Fixos



#### Fixos Headless Screw

- Specific Design Scarf & Phalangeal osteotomies (2.5 mm SV) Chevron osteotomies (3.5mm CS) Forefoot, Midfoot osteotomies and fusions (4.0mm MV)
- Titanium Alloy
- Different Thread Pitch between the head and the distal shaft to allow compression
- Cannulated
- Self Tapping/Self Drilling
- Optional Drill Countersink
- Hexagonal Screwdriver 1.8, 2.0, 2.5mm
- Guide Wire Diameter (0.9mm, 1mm, 1.4mm)



#### Fixos Twist-Off Screw

- Specific Design for Cancellous Anchorage Metatarsal Weil osteotomies of lesser rays (2.0mm WS) Chevron/Austinosteotomies (2.7mm KS)
- Titanium Alloy
- Self Tapping/Self Drilling
- Direct Power Insertion and final introduction with a specific screwdriver

**i** For more information on the Stryker Foot & Ankle Product Portfolio call today. 866.682.7577

A surgeon must always rely on his or her own professional clinical judgment when deciding whether to use a particular product when treating a particular patient. Stryker does not dispense medical advice and recommends that surgeons be trained in the use of any particular product before using it in surgery. The information presented is intended to demonstrate the breadth of Stryker product offerings. A surgeon must always refer to the package insert, product label and / or instructions for use before using any Stryker product. Products may not be available in all markets because product availability is subject to the regulatory and / or medical practices in individual markets. Please contact your Stryker representative if you have questions about the availability of Stryker products in your area. Stryker Corporation or its divisions or other corporate affiliated entities own, use or have applied for the following trademarks or service marks: Anchorage, Asnis, Fixos, Hoffmann, Stryker, T2, VariAx, Vitoss.

**Manufactured by:**  
Stryker Trauma AG  
Bohnackerweg 1  
CH - 2545 Selzach  
Switzerland  
www.osteosynthesis.stryker.com

**Distributed by:**  
Stryker  
325 Corporate Drive  
Mahwah, NJ 07430  
t: 201 831 5000  
www.stryker.com